



701R



Watson-Marlow Pumps Group
 Telephone +44 (0) 1326 370370
 Fax +44 (0) 1326 376009
 E-mail: info@watson-marlow.co.uk
 www.watson-marlow.com

Members of the Spirax Sarco Engineering Group



The information contained in this document is believed to be correct, but the Watson-Marlow Pumps Group accepts no liability for any errors it contains, and reserves the right to alter specifications without notice.



WARNING
 This product is not designed for use in, and should not be used for, patient-connected applications.

PB0311/10

	Max 316rpm		+5C, +41F
	<85dBA@1m		+40C, +104F
	701PB/R 20kg		-40C, -40F
			+70C, +158F



701PB/R



DS Cole 2006/42/EC
 David Cole, Managing director

